ACCU-CHEK® Insight

Roche

GET YOUR PUMP READY

INSERTING THE BATTERY

Make sure that the pump is **disconnected from your body**.

1

Push the battery cover release in the direction of the arrow.

2

Remove the battery cover.

3

Insert the new battery into the battery compartment, positive end first.

4

Push the battery cover all the way into the battery compartment until you hear a click.

Once the battery has been inserted, the pump begins its start-up process. If the pump does not begin the start-up process, make sure that the battery is inserted correctly (positive end first).

The first time you insert a battery or after the pump has been without battery for more than 10 hours, the Setup Wizard launches to help you program your pump's initial settings.

SETTING THE TIME, DATE AND BATTERY TYPE

1

Use ▲, ▼ and ◘ to set the time and date.

2

Set the battery type.

Make sure the correct battery type has been selected. Otherwise your pump may not display the W32: Battery low on time.

3

The **Start insulin delivery?** screen appears.
Press ◘ to select **no**.

4

The **Pump stopped** screen appears.

INSERTING A NEW CARTRIDGE

Make sure that the pump is **disconnected from your body**.

1

From the **Status screen**, press to reach the **Main Menu**.

2

Press ▼ to move to **Cartridge and infusion set**. Press to select.

3



The **Cartridge and infusion set menu** appears. Press to select **Change cartridge**.

4

The **Change cartridge** screen appears.

5

Hold the pump in an upright position. Turn the adapter with the tubing anti-clockwise.

6

When you reach an angle of 120º, remove the adapter with the cartridge. The pump starts rewinding the piston rod.

7

Disconnect infusion set from your body. Press to confirm.

8

Wait until the pump finishes rewinding the piston rod. If fully rewound this screen will not be displayed.

Before inserting the cartridge into the insulin pump, check every cartridge for cracks or damages. Never use a cartridge that has been dropped, even if it seems to be undamaged.

9

Insert a new cartridge only if the **Insert new cartridge and connect adapter** screen appears.

10

Make sure the pump is held in an upright position, insert the new cartridge - plunger first.

11

Place a **new adapter** on the bayonet socket of the cartridge compartment. Ensure to insert the nose of the adapter into the notches of the pumps housing at a right angle (90º). Otherwise the needle could bend.

12

Turn the adapter clockwise all the way until it stops.

13

When the adapter is correctly positioned, press to confirm.

14

The **Cartridge type** screen appears. Press to confirm.

15

The pump will automatically detect the plunger position.

16

The pump performs a self-test.

Continued over page

FILLING THE INFUSION SET TUBING AND CANNULA

1

The Fill tubing? screen appears after the self test. Press to fill the infusion set tubing with insulin.

2

Make sure the infusion set is not connected to your body. Press to confirm.

3

The pump starts to fill the tubing.

4

You can stop the filling process at any time by pressing . Filling is complete, when the pump has stopped. Check if insulin drips from the tubing.

5

Connect the tubing to the cannula. Press to confirm.

6

Press to fill the infusion set cannula with insulin.

7

The pump starts to fill the cannula.

8

The **Start insulin delivery?** screen appears. Press to select **yes**.

CHOOSING THE RIGHT ACCU-CHEK INSIGHT INFUSION SET

With the innovative needle guard systems and large “easy grip” handles, the Accu-Chek Insight infusion sets are designed for safety and ease of use in insulin pump therapy.

	Accu-Chek® Insight Flex	Accu-Chek® Insight Tender	Accu-Chek® Insight Rapid
Cannula type	Teflon		stainless steel
Insertion angle	90°	20°–45°	90°
Cannula length	6, 8 or 10 mm	13 or 17 mm	6, 8 or 10 mm
Tubing length	40, 70 or 100 cm		
Insertion process	manual or automatic	manual	

Important note:

Check your blood glucose 1–2 hours after changing your cannula and/or your cartridge. Avoid changing your cannula and cartridge immediately prior to going to sleep.

This quick start guide is not intended to replace the user manual, all users should read the user manual before use.

THE ACCU-CHEK PUMP CARELINE

24 hour telephone support for technical enquiries relating to the Accu-Chek Insight pump

TELEPHONE:
0800 731 2291 (UK), 1800 88 23 51 (ROI)

EMAIL:
burgesshill.insulinpumps@roche.com

WEBSITE:
www.accu-chek.co.uk

© 2020 Roche Diabetes Care Limited. All rights reserved.
ACCU-CHEK and ACCU-CHEK INSIGHT are trademarks of Roche.
All other trademarks or brand names are the property of their respective owners.

Roche Diabetes Care Limited, Charles Avenue, Burgess Hill, RH15 9RY, UK.
Company registration number: 9055599

For use in the UK and Ireland only
Date of preparation: September 2020
Material Number: 09886931001

www.accu-chek.co.uk **www.accu-chek.ie**

IDSCONS2020196