ACCU-CHEK® Insight

Roche

KEYS, MENU AND ESSENTIALS

- Dedicated, ergonomically designed adapter for an easy change of infusion set
- Very low Basal rate (0.02 U/h)
- Smart and sleek design for discretion with customisable screen (black or white)

THE PUMP MAIN MENU SCREEN

Information bar

Menu
Further functions can be accessed via ▼, e.g. **Start pump, Pump data**, etc.

THE PUMP STATUS SCREEN

Information bar

Pump status
Including information on mode, basal rate, basal profile, temporary basal rate or running bolus.

Continued over

THE KEYS

Up key

- Moves you up in a menu screen
- Increases a setting
- Function to access quick info screen

OK key

- Moves to the next menu level
- Confirms a setting
- Starts a function

Down key

- Moves you down in a menu screen
- Decreases a setting
- Function to access quick info screen

- Enlarges or reduces the characters on the display

Back/Quick bolus keys

Most often the 2 keys have an identical function and can be used alternatively:

- Moves to previous entry field
- Cancels a setting
- Stops a function
- Starts the **Quick bolus** programming
- Sets the **Quick bolus** amount
- Cancels a bolus (press and hold for 3 seconds)
- Moves to the **Status** screen (double click)
- Stops the repetitive STOP mode warning

Key lock disengages your pump's ▲, ▼ and ◉ keys so they cannot be pressed accidentally.

The **Key lock** is activated by default. It can be opened by pressing the following keys within 1 second:
one of the ⊂⊃ keys and then ▲ and then ▼ or vice versa.

You can adjust the time before key lock activates in **Settings** › **Device settings** › **Key lock**

TURNING FLIGHT MODE ON/OFF ON THE PUMP

1

From the **Status** screen, press ◉ to move to the **Menu screen.**

2

Press ▼ to move to **Modes.** Press ◉ to select.

3

Press ▼ to move to **Flight Mode.** Press ◉ to select.

4

Press ▲ or ▼ to move to **On** or **Off.** Press ◉ to confirm.

MODE SYMBOL – 3 BASIC MODES

The pump always displays the current mode by its respective symbol in the upper right corner of the screen.

Run

Your pump is delivering insulin according to the active basal profile and boluses programmed.

Pause

When you remove the battery while your pump is in RUN mode, your pump switches to PAUSE mode.

When the pump is in PAUSE mode for more than 15 min, it switches to STOP mode.

Stop

No insulin is being delivered, all boluses and TBR are cancelled. To re-start insulin delivery you have to switch to RUN mode using the Main Menu.
Boluses and TBR must be reprogrammed.

This quick start guide is not intended to replace the user manual, all users should read the user manual before use.

The Bluetooth® word mark and logos are registered trademarks owned by Bluetooth SIG, Inc. and any use of such marks by Roche is under license.

© 2020 Roche Diabetes Care Limited. All rights reserved.
ACCU-CHEK and ACCU-CHEK INSIGHT are trademarks of Roche.
All other trademarks or brand names are the property of their respective owners.

Roche Diabetes Care Limited, Charles Avenue, Burgess Hill, RH15 9RY, UK.
Company registration number: 9055599

For use in the UK and Ireland only
Date of preparation: September 2020.
Material Number: 09834575001

www.accu-chek.co.uk **www.accu-chek.ie**

IDSCONS2020195